

## NANO EXPRESS

## Open Access



# Preparation of Stabilizer-Free Silver Nanoparticle-Coated Micropipettes as Surface-Enhanced Raman Scattering Substrate for Single Cell Detection

Yi-bin Tan<sup>1</sup>, Jie-meng Zou<sup>1</sup> and Ning Gu<sup>1,2\*</sup>**Abstract**

In this work, we established a convenient while reproduceable method for stabilizer-free silver nanoparticle (AgNP)-coated micropipettes by the combination of magnetron sputtering and surface coupling agent. The clear surfaces of the AgNPs are beneficial for absorbing biological or functional molecules on their surfaces. By optimizing the operating parameters, such as sputtering current and sputtering time, the tip of micropipettes coated with AgNPs exhibits excellent surface-enhanced Raman scattering (SERS) performance. Finally, the Raman spectra of a single A549 lung adenocarcinoma cell are successfully acquired by these advanced SERS-active micropipettes.

**Keywords:** Micropipette; Magnetron sputtering; Silver nanoparticles; SERS; Single cell detection

**Background**

It is important to achieving detailed and accurate information of pathological process on cellular level in life science research. Generally, cell detection is based on indirect methods to collect certain components in a large amount of cells [1–3]. However, such statistically average results from a large amount of cells usually obscure the real mechanism of the cellular pathological process [4]. As a solution, detection of a single cell will help to comprehensively understand the individual difference and interaction [5, 6].

With the emerging of new technologies and instruments such as capillary electrophoresis, patch clamp, fluorescence microscope, and scanning tunneling microscope, the detection methods of a single cell have witnessed continuous development [7–10]. Raman spectroscopy, as a noninvasive optical detection method without ionizing radiation, has significant advantages like narrow spectrum peaks, interference immunity of water, hard quenching, and infrared light excitation compared with infrared spectroscopy and fluorescence spectroscopy, which is well

suited to the research of biosystem in solution [11–14]. However, since the ratio between a Raman scattering event and the incident photons upon a molecule is extremely low (~1 in 10 million), it is hard to obtain the Raman signals of a single cell. To conquer this problem, it is necessary to introduce nanostructured noble metals into the detection system. When the target molecules absorbed on the surface of these metal nanostructures, the amplification of the Raman scattering signals would be improved to several orders. This phenomenon is also called surface-enhanced Raman scattering (SERS) [15–17]. Some SERS-based methods have been exploited to enhance the Raman signals of single cells [18–20].

Micropipettes made of pulling glass capillary can be used in microinjection of a single cell [21]. Also, it can be used to observe the electrophysiological activities of single cells in patch clamp experiments [22]. Using such micropipettes with a metal nanostructures coating on its surface can serve as a kind of SERS-active microprobe for single cell detection. Since there is no need to introduce external markers when using this kind microprobe to conduct Raman detection, the information from the cell itself and its chemical response to the environment can be obtained accurately and conveniently [23–25]. Vitol et al. [26] utilized the electrostatic adsorption of citrate-stabilized gold nanoparticles on the surface of

\* Correspondence: [guning@seu.edu.cn](mailto:guning@seu.edu.cn)

<sup>1</sup>Jiangsu Key Laboratory for Biomaterials and Devices, State Key Laboratory of Bioelectronics, School of Biological Sciences and Medical Engineering, Southeast University, 2 Sipailou, Nanjing 210096, People's Republic of China

<sup>2</sup>Suzhou Key Lab of Biomedical Materials and Technology, Research Institute of Southeast University in Suzhou, 150 Ren Ai Road, Suzhou Industrial Park, Suzhou 215123, People's Republic of China

micropipette tips to fabricate SERS probes for detecting cell nucleus and cytoplasm.

In this paper, a facile while robust method is developed to obtain stabilizer-free silver nanoparticle (AgNP)-coated micro-SERS probes based on magnetron sputtering process [27, 28]. Furthermore, we investigated the effects of preparation process on the SERS performances of the probe tips. Finally, SERS detection for single cells was conducted using the as-prepared micropipettes.

## Methods

### Preparation of SERS Micropipettes

Initial micropipettes were prepared through standard wall borosilicate tubing (1.0 mm outer diameter, 0.5 mm inner diameter, Sutter Instruments, Novato, CA). The instrument used to pull the tubing is P-97 Flaming/Brown micropipette puller (Sutter Instruments, San Rafael, CA). Micropipettes with controllable taper length and tip size can be obtained by adjusting the parameters like pulling strength and speed (PULL and VEL). The pulled micropipettes were cleaned by acetone and deionized water, respectively, and dried for further use. In order to improve the deposition rate of the metal covering layer and enhance its conjugation with the substrate, 3-aminopropyltriethoxysilanes (APTS) were adopted as the coupling agent according to the silver. The metal covering layer on the surface of micropipettes was produced by magnetron sputtering method through Q150TS Sample Preparation System (Quorum Technologies Limited Company, Kent, UK),


which can operate on numerous micropipettes at the same time and realize mass production. The sputtering target material used in the system is silver (51 mm diameter, 1 mm thickness, 99.99 % purity, China New Metal Materials Technology, Beijing, China).

### Characterization Instruments

The surface morphology of micropipettes with silver covering layer was imaged by a field emission scanning electron microscope (SEM, Zeiss Ultra Plus, Germany) with InLens capabilities. To overcome the limitation of the working distance of the objective lens, we have built an inverted micro-Raman detection system (Fig. 1a) to aim the micropipette stuck into the sample conveniently. The whole system consists of excitation light source (532 nm single longitudinal mode solid laser, AOTK, Xiamen, China), inverted microscope (DMI3000B, Leica, Germany), micromanipulator (MP-225, Sutter Instrument, San Rafael, CA), Raman spectrometer (inVia series, 1800 lines/mm grating, Renishaw, UK), and the spectral analysis software (Wire3.4, Renishaw, UK).

### Raman Spectra Acquisition of Single Cells Using SERS Micropipettes

Nile Blue A (NBA, China National Medicine Corporation Ltd., Shanghai, China) was used as Raman molecule for evaluating the SERS performances of different micropipettes. The concentration of NBA is  $10^{-5}$  M and characteristic peak is  $592\text{ cm}^{-1}$ . The SERS micropipettes were immersed into the solution. The exciting



**Fig. 1** Raman detection system and single cell detection. **a** Diagrammatic drawing of the inverted microscope Raman detection system. **b** Photograph of the single cell Raman detection under optical microscopy with a 50x objective

source was focused on the sample through a  $\times 50$  telephoto objective lens to a spot size of approximately  $2\ \mu\text{m}$ . The Raman scattering was also collected by the same objective lens. The acquisition time for all spectra of micropipettes is 1 s, and the acquisition range is from  $300$  to  $1000\ \text{cm}^{-1}$ .


Monolayer cultures of A549 cells, a human lung adenocarcinoma cell line, grew in the 35-mm culture dish with Dulbecco's modified Eagle's medium (DMEM, KeyGEN BioTECH Company, Nanjing, China). These cells were supplemented with 10 % fetal bovine serum (FBS, SiJiQing Biomaterial Company, Hangzhou, China), and the temperature was maintained at  $37\ ^\circ\text{C}$  in a humidified, 5 % carbon dioxide atmosphere. A SERS-active micropipette is stuck into the cell (Fig. 1b). The acquisition time for all spectra of cell is 10 s, and the acquisition range is from  $400$  to  $2000\ \text{cm}^{-1}$ .

## Results and Discussion

### Raman Enhancement Influence of the Sputtering Parameters

In order to achieve great Raman enhancement, the plasmon resonance band of the SERS micropipettes should match the excitation wavelength [29]. It is directly related to the geometrical morphology of metal nanostructures on the substrate surface [30] and can be controlled by adjusting the sputtering parameters including sputtering current and sputtering time [31, 32].

Among these factors, sputtering current mainly affect the number of sputtering atoms and their mean kinetic energy. With the increased sputtering current, the mean kinetic energy of sputtering atoms will increase, which may promote the migration ability of sputtering atoms on the substrate surface. The combination between metal atoms and the substrate become more compact to form a smooth metal membrane, which is not preferable




**Fig. 2** Raman spectra of micropipettes measured with NBA ( $10^{-5}\text{M}$ ). **a** Micropipettes prepared by different sputtering currents. **b** Micropipettes prepared using different sputtering times


for the Raman enhancement. In this paper, we utilize different working currents and equal time to conduct the sputtering and choose the cylinder of the micropipettes to detect the SERS enhancement. By comparing the  $592\text{ cm}^{-1}$  peak intensity of Nile Blue A (Fig. 2a), the obtaining result agrees well with the description above.

The influence of sputtering time on SERS enhancement was quite different from that of the sputtering current [27, 33, 34]. At the initial stage of sputtering, only a few of critical nuclei are formed on the base


surface, and the size of metal particles is small. Here, the distance between particles is too large to enhance Raman signals by electromagnetic mechanism. When the distance between these metal particles reached the critical value to form electromagnetic enhancement, SERS signal appears and presents an increasing trend. When the metal nanostructure on the surface matched with the wavelength of the excitation light, the SERS enhancement achieves the peak value. By doing this, the surface of the substrate usually shows rough island-like



**Fig. 3** Characterization of the SERS micropipettes in different sections. **a** Photograph of optical microscopy and SEM. Three sections (A, B, C) of the micropipette with different curvatures are chosen for the observation of the film morphologies. The corresponding curvatures of sections A, B, and C are listed in the table. **b** Raman spectra of SERS micropipette with the detection location moved from the tip to the column end



**Fig. 4** Characterization of the SERS micropipettes with different tip angles. **a** Comparison of the normalized peak intensity at 592 cm<sup>-1</sup> from the tip to the column end on micropipettes. **b** The process schematic of formation of the metal membrane on micropipettes




**Fig. 5** Characterization of SERS micropipettes prepared with or without coupling agent in different sputtering conditions

membrane or network membrane structure. As the process of sputtering, continuous membrane started to form, and the roughness of substrate surface decreased, and the SERS enhancement would decrease gradually again (Fig. 2b).

#### Raman Enhancement Influence of the Tip Geometry

For conical micropipette tips, in addition to the sputtering parameters, their own geometry also has great impact on the morphology of the metal covering layer. The tip of a SERS micropipette with a minimal diameter of about 2.5  $\mu\text{m}$  is characterized under optical microscopy

and SEM as shown in Fig. 3a. Three sections of the micropipette with different curvatures are chosen for the observation. Similar to the process of membrane forming with different sputtering times, critical nuclei, rough island-like membrane, and network membrane structure are found in the different sections. These phenomena show that the deposition of sputtering atoms at different sections of conical micropipette is determined by the curvature. It is more hardful for sputtering atoms attaching on the tip with large curvature than that on the column end. With the decrease of surface curvature, the formation of continuous film can be more quickly. SERS



**Fig. 6** A549 cell detection with or without the SERS micropipettes. **a** Comparison of mean Raman spectra. **b** Comparison of the difference SERS spectra



spectra of different sections also endorse this view. As the detection location moved from the tip to the column end, the SERS enhancement of micropipette presented an increasing-to-decreasing trend (Fig. 3b).

Figure 4a shows the experimental results of similar SERS detection for a group of micropipettes with different taper angles but similar tip size. The result clearly indicated that, for single micropipette, the changing trend of characteristic  $592\text{ cm}^{-1}$  peak intensity was the same as the result indicated in Fig. 3b. Seen from the comparison between tips with different angles, as the angle decreased, the changing speed of curvature from the tip to the column end was slowing down, and the location where the maximum  $592\text{ cm}^{-1}$  characteristic peak intensity is increasingly far from the tip. The process schematic of formation of the metal membrane on micropipettes with different tip angles is given by Fig. 4b.

The above study and analysis were of vital importance in actual preparation of micropipettes. If micropipettes have better SERS enhancement, it would be quite helpful for obtaining weaker Raman signal at a single cell level. To achieve this goal, it is required to use different preparation parameters for micropipettes with different tip sizes. For this regard, we will fix the sputtering current and adjust the sputtering time to realize this goal.

### The Influence of Coupling Agent

Without coupling treatment, the metal atoms generated by magnetron sputtering would attach to the substrate surface by physical absorption. As a result, it was not tight enough so that it is easy to be affected by the environment. When APTS is used as the coupling agent, monomolecular layer with regular shape is formed on the surface of micropipettes under waterless condition, thus being helpful for the attachment of sputtering atoms. To enhance the interaction between covering metal layer and substrate, a method of accelerating the deposition speed of sputtering atoms, reducing the sputtering time, and the wastage of target materials is used. Fig. 5 demonstrates the mean  $592\text{ cm}^{-1}$  peak intensity of five groups of SERS micropipettes prepared with and without coupling agents. The result indicated that, with the presence of coupling agent, only a short sputtering time was required to obtain micropipettes with good SERS enhancement.

### Experimental Result of Single Cell Detection

We apply the prepared SERS micropipettes to conduct Raman spectrum detection on A549 cells. As shown in Fig. 6a, the cell culture environment such as culture dish and culture medium also have strong Raman signal, especially at  $620, 795, 1001, 1031, 1154, 1197, 1448$ , and  $1602\text{ cm}^{-1}$  due to the biomolecule vibration, which would overlap the required information. Without SERS

micropipettes, it is very difficult to obtain the weak Raman signal of A549 cells grew closely attaching to the bottom of the culture dish, even by confocal microscopy. Indeed, the Raman spectra of A549 cells without the presence of SERS micropipettes are nearly the same as the controlled spectra of the dish and medium (Fig. 6a). When SERS-active micropipettes are introduced, it can be clearly discovered that abundant characteristic peak signals attributed to the A549 cell including  $756, 938, 1249, 1363$ , and  $1659\text{ cm}^{-1}$  [35]. The above results clearly indicated that Raman signal can be enhanced at the tangent location of cell and micropipette. Through differential spectrum analysis and normalization processing, we can obtain richly spectroscopic information of the cells (Fig. 6b).

### Conclusions

In summary, we demonstrated a robust method of stabilizer-free AgNP-coated micro-SERS probes by magnetron sputtering and surface coupling modification. Using these SERS-active micro-substrate, we successfully obtain rich Raman signals of single cells in situ. This mass-production method owns a simple process, fast speed, and good repeatability. By adjusting the preparation parameters, we can optimize the SERS enhancement on the micropipettes and improve its detection sensibility. In future work, we will address the application of these SERS micropipettes for cell microinjection and electrophysiological detection, which is expected to realize a multi-parameter detection of single cells in situ.

### Competing Interests

The authors declare that they have no competing interests.

### Authors' Contributions

YBT carried out the sample preparation, system construction, spectrum measurements, and drafted the manuscript. YBT and JMZ did the data analysis and interpreted the results. JMZ helped to finish SEM sample preparation and observation. NG supervised the whole work and revised the manuscript. All authors read and approved the final manuscript.

### Acknowledgements

This work was supported by the National Basic Research Program of China (2011CB933503), the Special Project on Development of National Key Scientific Instruments Equipment of China (2011YQ03013403), Special Funds of the National Natural Science Foundation of China for Basic Research Projects of Scientific Instruments (61127002), National Natural Science Foundation of China (61179035), the National Natural Science Foundation of China for Key Project of International Cooperation (61420106012), and Collaborative Innovation Center of Suzhou Nano Science and Technology.

Received: 6 June 2015 Accepted: 15 October 2015

Published online: 26 October 2015

### References

- Gerlier D, Thomasset N (1986) Use of MTT colorimetric assay to measure cell activation. *J Immunol Methods* 94(1):57–63
- Karube I, Nakahara T, Matsunaga T, Suzuki S (1982) Salmonella electrode for screening mutagens. *Anal Chem* 54(11):1725–7
- Li X, Schwartz RM, Cesar EY, Wang HY (1988) An integrated microcomputer system using immobilized cellular electrodes for drug screening. *Comput Biol Med* 18(5):367–76

4. Carlo DD, Lee LP (2006) Dynamic single-cell analysis for quantitative biology. *Anal Chem* 78(23):7918–25
5. Lu X, Huang W, Wang Z, Cheng J (2004) Recent developments in single-cell analysis. *Anal Chim Acta* 510(2):127–38
6. Cross SE, Jin Y, Rao J, Gimzewski JK (2007) Nanomechanical analysis of cells from cancer patients. *Nat Nanotechnol* 2(12):780–3
7. Schulte A, Schuhmann W (2007) Single-cell microelectrochemistry. *Angew Chem Int Ed* 46(46):8760–77
8. Huang Y, Cai D, Chen P (2011) Micro-and nanotechnologies for study of cell secretion. *Anal Chem* 83(12):4393–406
9. Wang X, Veruki ML, Bukoreshtliev NV, Hartveit E, Gerdes H (2010) Animal cells connected by nanotubes can be electrically coupled through interposed gap-junction channels. *Proc Natl Acad Sci* 107(40):17194–9
10. Sample V, Newman RH, Zhang J (2009) The structure and function of fluorescent proteins. *Chem Soc Rev* 38(10):2852–64
11. Xie C, Li Y (2003) Confocal micro-Raman spectroscopy of single biological cells using optical trapping and shifted excitation difference techniques. *J Appl Phys* 93(5):2982–6
12. Xie C, Dinno MA, Li Y (2002) Near-infrared Raman spectroscopy of single optically trapped biological cells. *Opt Lett* 27(4):249–51
13. Baraga JJ, Feld MS, Rava RP (1992) Rapid near-infrared Raman spectroscopy of human tissue with a spectrograph and CCD detector. *Appl Spectrosc* 46(2):187–90
14. Schie IW, Huser T (2013) Methods and applications of Raman microspectroscopy to single-cell analysis. *Appl Spectrosc* 67(8):813–28
15. Nie S, Emory SR (1997) Probing single molecules and single nanoparticles by surface-enhanced Raman scattering. *Science* 275(5303):1102–6
16. Kneipp K, Wang Y, Kneipp H, Perelman LT, Itzkan I, Dasari RR, Feld MS (1997) Single molecule detection using surface-enhanced Raman scattering (SERS). *Phys Rev Lett* 78(9):1667
17. Lane LA, Qian X, Nie S (2015) SERS nanoparticles in medicine: from label-free detection to spectroscopic tagging. *Chem Rev* 115(19):10489–529. doi:10.1021/acs.chemrev.5b00265
18. Kneipp K, Haka AS, Kneipp H, Badizadegan K, Yoshizawa N, Boone C, Shafer-Peltier KE, Motz JT, Dasari RR, Feld MS (2002) Surface-enhanced Raman spectroscopy in single living cells using gold nanoparticles. *Appl Spectrosc* 56(2):150–4
19. Palonpon AF, Ando J, Yamakoshi H, Dodo K, Sodeoka M, Kawata S, Fujita K (2013) Raman and SERS microscopy for molecular imaging of live cells. *Nat Protoc* 8(4):677–92
20. Bankapur A, Krishnamurthy RS, Zachariah E, Santhosh C, Chougule B, Praveen B, Valiathan M, Mathur D (2012) Micro-Raman spectroscopy of silver nanoparticle induced stress on optically-trapped stem cells. *Plos One* 7(4):e35075
21. Capecchi MR (1980) High efficiency transformation by direct microinjection of DNA into cultured mammalian cells. *Cell* 22(2):479–88
22. Hamill OP, Marty A, Neher E, Sakmann B, Sigworth FJ (1981) Improved patch-clamp techniques for high-resolution current recording from cells and cell-free membrane patches. *Pflugers Arch* 391(2):85–100
23. Vitol EA, Orynbayeva Z, Friedman G, Gogotsi Y (2012) Nanoprobes for intracellular and single cell surface-enhanced Raman spectroscopy (SERS). *J Raman Spectrosc* 43(7):817–27
24. Cullum BM, Vo-Dinh T (2000) The development of optical nanosensors for biological measurements. *Trends Biotechnol* 18(9):388–93
25. Scaffidi JP, Gregas MK, Seewaldt V, Vo-Dinh T (2009) SERS-based plasmonic nanobiosensing in single living cells. *Anal Bioanal Chem* 393(4):1135–41
26. Vitol EA, Orynbayeva Z, Bouchard MJ, Azizkhan-Clifford J, Friedman G, Gogotsi Y (2009) In situ intracellular spectroscopy with surface enhanced Raman spectroscopy (SERS)-enabled nanopipettes. *ACS Nano* 3(11):3529–36
27. Kelly PJ, Arnell RD (2000) Magnetron sputtering: a review of recent developments and applications. *Vacuum* 56(3):159–72
28. Biederman H (2000) RF sputtering of polymers and its potential application. *Vacuum* 59(2–3):594–9
29. Moskovits M (1985) Surface-enhanced spectroscopy. *Rev Mod Phys* 57(3):783
30. Shirlcliffe N, Nickel U, Schneider S (1999) Reproducible preparation of silver sols with small particle size using borohydride reduction: for use as nuclei for preparation of larger particles. *J Colloid Interface Sci* 211(1):122–9
31. Kumru M (1991) A comparison of the optical, IR, electron spin resonance and conductivity properties of a-Ge<sub>1-x</sub>C<sub>x</sub>H with a-Ge:H and a-Ge thin films prepared by rf sputtering. *Thin Solid Films* 198(1):75–84
32. Anders A, Andersson J, Ehasarian A (2007) High power impulse magnetron sputtering: current-voltage-time characteristics indicate the onset of sustained self-sputtering. *J Appl Phys* 102(11):113303
33. Jung YS (2004) Study on texture evolution and properties of silver thin films prepared by sputtering deposition. *Appl Surf Sci* 221(1–4):281–7
34. Sun X, Hong R, Hou H, Fan Z, Shao J (2007) Thickness dependence of structure and optical properties of silver films deposited by magnetron sputtering. *Thin Solid Films* 515(17):6962–6
35. Wang M, Cao X, Lu W, Tao L, Zhao H, Wang Y, Guo M, Dong J, Qian W (2014) Surface-enhanced Raman spectroscopic detection and differentiation of lung cancer cell lines (A549, H1229) and normal cell line (AT II) based on gold nanostar substrates. *RSC Adv* 4(19):64225–34

**Submit your manuscript to a SpringerOpen<sup>®</sup> journal and benefit from:**

- Convenient online submission
- Rigorous peer review
- Immediate publication on acceptance
- Open access: articles freely available online
- High visibility within the field
- Retaining the copyright to your article

Submit your next manuscript at ► [springeropen.com](http://springeropen.com)